

**Notified Body Confirmation Letter**

2024-03-26

DELMONT IMAGING
390 Avenue du Mistral

FR-13600 LA CIOTAT

Notified Body Confirmation Letter**Reference: 5810-2024/01**

To whom it may concern,

Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 as regards the transitional provisions for certain medical devices.

This letter confirms that, SIQ Ljubljana, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number 1304 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

Company Name	DELMONT IMAGING
Legal address/street	390 Avenue du Mistral
Zip code/town	FR-13600 LA CIOTAT
Country:	France
SRN number	FR-MF-000000157

The devices covered by the formal application and the written agreement mentioned above are identified in the Tables below. Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive. Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but the NB has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.

In the case of devices covered by certificates issued under Directive 93/42/EEC (MDD) that expired after 26 May 2021 and before 20 March 2023, without having been withdrawn, this letter also confirms that the manufacturer signed the written agreement under MDR by the date of MDD certificate expiry; or provided evidence that a competent authority of a Member

Notified Body Confirmation Letter

State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively, by the 20 Mar 2023 for the relevant devices.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3c of MDR (as amended by (EU) 2023/607), are shown below:

- 26 May 2026 for Class III custom-made implantable devices;
- 31 December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors);
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function;
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments).

On behalf of the Notified Body,

Ana Pribaković Borštnik
Product manager MDR



Table 1: Devices covered by this letter and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:

DEVICE NAME OR BASIC UDI-DI (under MDR Application)	MDR DEVICE CLASSIFICATION (as proposed by the manufacturer and verified at the pre-application stage)	IF THE MDR DEVICE IS A SUBSTITUTE DEVICE, IDENTIFICATION OF THE CORRESPONDING MDD DEVICE	MDD CERTIFICATE REFERENCE(S) OF THE DEVICES UNDER MDR APPLICATION, AND THE NB IDENTIFICATION
Resectoscopy System 37012178RESSLA	Class IIb	N/A	Certificate MDD-085; NB1304
Rigid Endoscopes 37012178ENDOH7	Class IIa	N/A	Certificate MDD-085; NB1304
Hysteroscopy System 37012178HYSSM8	Class IIb	N/A	Certificate MDD-085; NB1304
Minimally invasive HF instrument 37012178HFINGZ	Class IIb	N/A	Certificate MDD-085; NB1304
Uterine manipulator 37012178UTMAML	Class IIa	N/A	Certificate MDD-085; NB1304

Notified Body Confirmation Letter

Table 2: Devices covered by this letter and for which the NB is NOT responsible for appropriate surveillance of the corresponding devices under the applicable Directive

DEVICE NAME OR BASIC UDI-DI (under MDR Application)	MDR DEVICE CLASSIFICATION (as proposed by the manufacturer and verified at the pre-application stage)	IF THE MDR DEVICE IS A SUBSTITUTE DEVICE, IDENTIFICATION OF THE CORRESPONDING MDD DEVICE	MDD CERTIFICATE REFERENCE(S) OF THE DEVICES UNDER MDR APPLICATION, AND THE NB IDENTIFICATION
Irrigation fluid management system 37012178IFLMHF	Class IIb	N/A	N/A
Electrosurgical unit for general use 37012178ESUGK3	Class IIb	N/A	N/A

Confirmation Letter Revision History

DATE	NB INTERNAL REFERENCE TRACEABLE TO EACH VERSION OF THE LETTER	ACTION
2024-03-26	5810-2024/ 01	Initial issue